

## Idaho Bureau of Laboratories

2220 Old Penitentiary Road  
Boise, Idaho 83712  
Phone: (208)334-2235  
Fax: (208)334-4067

Christopher L. Ball, Ph.D, Bureau Chief  
EPA Lab ID00018  
CLIA Certified Lab

October 13, 2010

**Report To** Todd Vitek  
V4 Distributing  
507 W Bird Ave  
Nampa, ID 83686

**Sample Submitted by** Walk-In Client - Private

RE: Workorder E100700201 072710  
Profile WA-General Environmental

Enclosed are the analytical results for sample(s) received by the laboratory on Tuesday, July 27, 2010. Results reported herein conform to the most current federal standards, where applicable, unless otherwise narrated in the body of the report.

If you have any questions concerning this report, please feel free to contact me.


Sincerely,

Ernie Bader

Enclosures

### CERTIFICATE OF ANALYSIS

This report shall not be reproduced, except in full,  
without the written consent of Idaho Bureau of Laboratories.



## Idaho Bureau of Laboratories

2220 Old Penitentiary Road  
Boise, Idaho 83712  
Phone: (208)334-2235  
Fax: (208)334-4067

Christopher L. Ball, Ph.D, Bureau Chief  
EPA Lab ID00018  
CLIA Certified Lab

Workorder: E100700201 072710

### SAMPLE SUMMARY

---

Lab ID	Sample ID	Matrix	Date Collected	Date Received
E100700201-001	Safe Kleen	Water		7/27/2010 11:07

---

### CERTIFICATE OF ANALYSIS

This report shall not be reproduced, except in full,  
without the written consent of Idaho Bureau of Laboratories.



# Idaho Bureau of Laboratories

2220 Old Penitentiary Road  
Boise, Idaho 83712  
Phone: (208)334-2235  
Fax: (208)334-4067

Christopher L. Ball, Ph.D, Bureau Chief  
EPA Lab ID00018  
CLIA Certified Lab

Workorder: E100700201 072710

## ANALYTICAL RESULTS

**Chemistry samples are disposed of after 60 days unless the laboratory is notified otherwise.**

Lab ID: **E100700201-001**      Date Received: 7/27/2010 11:07      Matrix: Water  
Sample ID: **Safe Kleen**      Date Collected:  
Location: **V4 Distributing**

Parameters	Results Units	Report Limit	Analyzed	By	Qual	MCL
Analysis Desc: Acode for test w/o Acode		Analytical Method: Acode for test w/o Acode				
Cyanide	Not Detected		9/3/2010	EB	3,2	
Alpha-BHC by EPA 608	< 100ug/L		8/19/2010	RA		
Gamma-BHC by EPA 608	< 100ug/L		8/19/2010	RA		
Beta-BHC by EPA 608	< 100ug/L		8/19/2010	RA		
Delta-BHC by EPA 608	< 100ug/L		8/19/2010	RA		
Heptachlor by EPA 608	< 100ug/L		8/19/2010	RA		
Aldrin by EPA 608	< 100ug/L		8/19/2010	RA		
Heptachlor Epoxide-EPA 608	< 100ug/L		8/19/2010	RA		
DDE by EPA 608	< 100ug/L		8/19/2010	RA		
Endosulfan I by EPA 608	< 100ug/L		8/19/2010	RA		
Dieldrin by EPA 608	< 100ug/L		8/19/2010	RA		
Endrin by EPA 608	< 100ug/L		8/19/2010	RA		
DDD by EPA 608	< 100ug/L		8/19/2010	RA		
Endosulfan II by EPA 608	< 100ug/L		8/19/2010	RA		
DDT & Endrin Ketone-EPA 608	< 100ug/L		8/19/2010	RA		
Endrin Aldehyde by EPA 608	< 100ug/L		8/19/2010	RA		
Methoxychlor by EPA 608	< 100ug/L		8/19/2010	RA		
Endosulfan Sulfate by EPA 608	< 100ug/L		8/19/2010	RA		
Chlordane by EPA 608	< 1000ug/L		8/19/2010	RA		
Toxaphene by EPA 608	< 1000 ug/L		8/19/2010	RA		
Arochlors by EPA 608	< 1000ug/L		8/19/2010	RA		

-- Volatiles --

Report ID: 79011 - 987037

Page 3 of 7

### CERTIFICATE OF ANALYSIS

This report shall not be reproduced, except in full, without the written consent of Idaho Bureau of Laboratories.



# Idaho Bureau of Laboratories

2220 Old Penitentiary Road  
Boise, Idaho 83712  
Phone: (208)334-2235  
Fax: (208)334-4067

Christopher L. Ball, Ph.D, Bureau Chief  
EPA Lab ID00018  
CLIA Certified Lab

Workorder: E100700201 072710

## ANALYTICAL RESULTS

**Chemistry samples are disposed of after 60 days unless the laboratory is notified otherwise.**

Lab ID: **E100700201-001** Date Received: 7/27/2010 11:07 Matrix: Water  
Sample ID: **Safe Kleen** Date Collected:  
Location: **V4 Distributing**

Parameters	Results	Units	Report Limit	Analyzed	By	Qual	MCL
Analysis Desc: EPA 624		Analytical Method: EPA 624					
Dichlorodifluoromethane	<120	ug/l	120	8/9/2010	RA		
Chloromethane	<120	ug/l	120	8/9/2010	RA		
Vinyl chloride	<120	ug/l	120	8/9/2010	RA		
Bromomethane	<120	ug/l	120	8/9/2010	RA		
Chloroethane	<120	ug/l	120	8/9/2010	RA		
Trichlorofluoromethane	<120	ug/l	120	8/9/2010	RA		
1,1-Dichloroethene	<120	ug/l	120	8/9/2010	RA		
Methylene chloride	<120	ug/l	120	8/9/2010	RA		
trans-1,2-Dichloroethene	<120	ug/l	120	8/9/2010	RA		
1,1-Dichloroethane	<120	ug/l	120	8/9/2010	RA		
2,2-Dichloropropane	<120	ug/l	120	8/9/2010	RA		
cis-1,2-Dichloroethene	<120	ug/l	120	8/9/2010	RA		
Bromochloromethane	<120	ug/l	120	8/9/2010	RA		
Chloroform	<120	ug/l	120	8/9/2010	RA		
1,1,1-Trichloroethane	<120	ug/l	120	8/9/2010	RA		
Carbon tetrachloride	<120	ug/l	120	8/9/2010	RA		
1,1-Dichloropropene	<120	ug/l	120	8/9/2010	RA		
Benzene	<120	ug/l	120	8/9/2010	RA		
1,2-Dichloroethane	<120	ug/l	120	8/9/2010	RA		
Trichloroethene	<120	ug/l	120	8/9/2010	RA		
1,2-Dichloropropane	<120	ug/l	120	8/9/2010	RA		
Dibromomethane	<120	ug/l	120	8/9/2010	RA		
Bromodichloromethane	<120	ug/l	120	8/9/2010	RA		
cis-1,3-Dichloropropene	<120	ug/l	120	8/9/2010	RA		
Toluene	<120	ug/l	120	8/9/2010	RA		
trans-1,3-Dichloropropene	<120	ug/l	120	8/9/2010	RA		
Tetrachloroethene	<120	ug/l	120	8/9/2010	RA		
1,3-Dichloropropane	<120	ug/l	120	8/9/2010	RA		

### CERTIFICATE OF ANALYSIS

This report shall not be reproduced, except in full, without the written consent of Idaho Bureau of Laboratories.



# Idaho Bureau of Laboratories

2220 Old Penitentiary Road  
Boise, Idaho 83712  
Phone: (208)334-2235  
Fax: (208)334-4067

Christopher L. Ball, Ph.D, Bureau Chief  
EPA Lab ID00018  
CLIA Certified Lab

Workorder: E100700201 072710

## ANALYTICAL RESULTS


**Chemistry samples are disposed of after 60 days unless the laboratory is notified otherwise.**

Lab ID: **E100700201-001**      Date Received: 7/27/2010 11:07      Matrix: Water  
Sample ID: **Safe Kleen**      Date Collected:  
Location: **V4 Distributing**

Dibromochloromethane	<120 ug/l	120	8/9/2010	RA
1,2-Dibromoethane	<120 ug/l	120	8/9/2010	RA
Chlorobenzene	<120 ug/l	120	8/9/2010	RA
Ethylbenzene	<120 ug/l	120	8/9/2010	RA
1,1,1,2-Tetrachloroethane	<120 ug/l	120	8/9/2010	RA
m,p-Xylene	<250 ug/l	250	8/9/2010	RA
o-Xylene	<120 ug/l	120	8/9/2010	RA
Styrene	<120 ug/l	120	8/9/2010	RA
Bromoform	<120 ug/l	120	8/9/2010	RA
Isopropylbenzene (Cumene)	<120 ug/l	120	8/9/2010	RA
Bromobenzene	<120 ug/l	120	8/9/2010	RA
1,1,2,2-Tetrachloroethane	<120 ug/l	120	8/9/2010	RA
n-Propylbenzene	<120 ug/l	120	8/9/2010	RA
1,2,3-Trichloropropane	<120 ug/l	120	8/9/2010	RA
2-Chlorotoluene	<120 ug/l	120	8/9/2010	RA
1,3,5-Trimethylbenzene	<120 ug/l	120	8/9/2010	RA
4-Chlorotoluene	<120 ug/l	120	8/9/2010	RA
tert-Butylbenzene	<120 ug/l	120	8/9/2010	RA
1,2,4-Trimethylbenzene	<120 ug/l	120	8/9/2010	RA
sec-Butylbenzene	<120 ug/l	120	8/9/2010	RA
p-Isopropyltoluene	<120 ug/l	120	8/9/2010	RA
1,3-Dichlorobenzene	<120 ug/l	120	8/9/2010	RA
1,4-Dichlorobenzene	<120 ug/l	120	8/9/2010	RA
n-Butylbenzene	<120 ug/l	120	8/9/2010	RA
1,2-Dichlorobenzene	<120 ug/l	120	8/9/2010	RA
1,2-Dibromo-3-chloropropane	<120 ug/l	120	8/9/2010	RA
1,2,4-Trichlorobenzene	<120 ug/l	120	8/9/2010	RA
Hexachlorobutadiene	<120 ug/l	120	8/9/2010	RA

### CERTIFICATE OF ANALYSIS

This report shall not be reproduced, except in full, without the written consent of Idaho Bureau of Laboratories.



# Idaho Bureau of Laboratories

2220 Old Penitentiary Road  
Boise, Idaho 83712  
Phone: (208)334-2235  
Fax: (208)334-4067

Christopher L. Ball, Ph.D, Bureau Chief  
EPA Lab ID00018  
CLIA Certified Lab

Workorder: E100700201 072710

## ANALYTICAL RESULTS

Chemistry samples are disposed of after 60 days unless the laboratory is notified otherwise.

Lab ID: **E100700201-001** Date Received: 7/27/2010 11:07 Matrix: Water  
Sample ID: **Safe Kleen** Date Collected:  
Location: **V4 Distributing**


Naphthalene	<120 ug/l	120	8/9/2010	RA	1
1,2,3-Trichlorobenzene	<120 ug/l	120	8/9/2010	RA	
Tent Id Cmpd	1,1'-oxybis-pentane		8/9/2010	RA	
Tent Id Cmpd	4-Nonanone		8/9/2010	RA	
Tent Id Cmpd	2-Nonen-1-ol		8/9/2010	RA	
Tent Id Cmpd	5-methyl-2-(1-methylethyl)-1-hexanol		8/9/2010	RA	
Tent Id Cmpd	2-propyl-1-heptanol		8/9/2010	RA	
Bromofluoromethane (S)	25200 %	-	8/9/2010	RA	
1,2-Dichloroethane-d4 (S)	24200 %	-	8/9/2010	RA	
Toluene-d8 (S)	26200 %	-	8/9/2010	RA	
Bromofluorobenzene (S)	24800 %	-	8/9/2010	RA	

### -- Metals --

Parameters	Results	Units	Report Limit	Analyzed	By	Qual	MCL
Analysis Desc: EPA 200.8		Analytical Method: EPA 200.8					
Arsenic	<0.31	mg/L	0.31	10/8/2010	HH		
Cadmium	<0.16	mg/L	0.16	10/8/2010	HH		
Chromium	<0.16	mg/L	0.16	10/8/2010	HH		
Copper	<0.16	mg/L	0.16	10/8/2010	HH		
Lead	<0.16	mg/L	0.16	10/8/2010	HH		
Nickel	<0.16	mg/L	0.16	10/8/2010	HH		
Zinc	<0.16	mg/L	0.16	10/8/2010	HH		
Analysis Desc: EPA 245.1		Preparation Method: EPA 245.1					
		Analytical Method: EPA 245.1					
Mercury	<0.025	mg/L	0.025	8/31/2010	JS		

### CERTIFICATE OF ANALYSIS

This report shall not be reproduced, except in full, without the written consent of Idaho Bureau of Laboratories.



## Idaho Bureau of Laboratories

2220 Old Penitentiary Road

Boise, Idaho 83712

Phone: (208)334-2235

Fax: (208)334-4067

Christopher L. Ball, Ph.D, Bureau Chief

EPA Lab ID00018

CLIA Certified Lab

Workorder: E100700201 072710

### ANALYTICAL RESULTS QUALIFIERS

---

#### PARAMETER QUALIFIERS

- [1] Sample run at 1:250 dilution due to excessive foaming
- [2] Cyanide Reporting Limit= 25 ug/L
- [3] Dissolved Cyanide analyzed using GC/MS Headspace method

#### CERTIFICATE OF ANALYSIS

This report shall not be reproduced, except in full,  
without the written consent of Idaho Bureau of Laboratories.